

Certificate of Analysis

Product Compact Dry SL for Salmonella
ID-No. 1 002 973 / 1 002 938 / 1 002 940
1 402 938 / 1 402 887 / 1 502 887
Lot Number 265106
Expiry Date 30.11.2022

Test	Specifications	Results*
Appearance	Purple blue colour	Corresponds
Loss on drying	< 10,0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	5.5 – 6.3	Corresponds
Colony growth	Color development	
<i>Salmonella typhimurium</i> ATCC 13311	yellow	Corresponds
<i>Salmonella enteritidis</i> SAL 1	yellow	Corresponds
<i>Salmonella habana</i> SAL 23-150	yellow	Corresponds
<i>Salmonella cubana</i> SAL FDA H ₂ S	yellow	Corresponds
<i>Citrobacter freundii</i> ATCC 8090	no growth	
<i>Escherichia coli</i> ATCC 8739	no growth	
<i>Klebsiella pneumoniae</i> ATCC 13880	no growth	
<i>Proteus mirabilis</i> ATCC 29906	no growth	
<i>Staphylococcus aureus</i> ATCC 6538	no growth	

This lot meets the specifications.

Date: 01.10.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.