

Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code:60361

Lot No.: 20210305Y

Expiry date: Jun. 04, 2022

TEST PARAMETER

RESULT

SPECIFICATION

When applied 100μL of 5pmol ADP
Average Luminescence(RLU)

Nen applied 100μL of 5pmol AMP
Average Luminescence(RLU)

12300

≥9100, ≤16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Ryosuke Sugimoto: Manager

Date Mar. 18, 2021

KBG6036100ac-Eaa SA2103-0179 Quality Assurance Group, Edogawa Quality Unit Kikkoman Biochemifa Company