

**CERTIFICATE OF ANALYSIS****LuciPac A3 Surface**

Product code :60361


Lot No.: 20220613Y

Expiry date: Sep. 12, 2023

| <u>TEST PARAMETER</u>                                              | <u>RESULT</u> | <u>SPECIFICATION</u>    |
|--------------------------------------------------------------------|---------------|-------------------------|
| When applied 100 $\mu$ L of 5pmol ADP<br>Average Luminescence(RLU) | 12600         | $\geq 9100, \leq 16900$ |
| When applied 100 $\mu$ L of 5pmol AMP<br>Average Luminescence(RLU) | 12400         | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
\_\_\_\_\_  
Ryosuke Sugimoto; ManagerDate Jun. 22, 2022KBG6036100ac-Eaa  
SA2206-0211Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company